Fox Valley Metrology

3114 Medalist Drive Oshkosh, WI 54902 (920) 426-5894 • Fax (920) 426-8120 www.FoxValleyMetrology.com

CERTIFICATE OF CALIBRATION




Certification Number CI064-38579-511		Procedures Followed FVE-015 rev. 3		This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results		
					relate only to the item(s) calibrated Form Revision 6: 02/04/2012	
Purchase Order #						
Test Instrument	Surface Plate (Small) 9x12, Grade: A		Standards Used Instrument	Serial Number	Trace Number	Next Cal
Make	Standard Surface Plates		FVR-173	N/A	CH158-26227-397	06/28/2018
Model	Small		FVS-582	11100244	CH243-41017-328	08/28/2018
Serial Number						
Identification	201703667					
Customer Location						
Condition Received	In Tolerance		Total expanded meas (k=2) is Repeatability	,	v based on two standard ness (66 + 0.2D) uin	deviations
Condition Received	In Tolerance					
Calibrated By	Scott Showers		All instruments have been calibrated against standards traceable to NIST. Calibration was completed in accordance with ISO 6789, ISO/IEC 17025:2005			
Form Reviewed By	M.T.		and ANSI Z540-1-1994.			
Calibration Location						
Calibration Conditions	59.0°F, 15.0°C, 26.0%RH					
Calibration Date	03/05/2018					
Recalibration Due	03/05/2019					

Calibration Results

denotes "Out of Tolerance"

±XX - denotes As Found **±XX** - denotes As Left



Comments

Size: 9x12 | Grade: A | Tolerance: 100 µin | Total Flatness: 100 µin | All readings are in microinches (µin).